

A Motion Correction Approach for Oral and Maxillofacial Cone-beam CT Imaging

Tao Sun¹, Reinhilde Jacobs², Ruben Pauwels³,
Elisabeth Tijskens², Roger Fulton^{4,5} and Johan Nuyts⁶

¹Paul C. Lauterbur Research Center for Biomedical Imaging, Shenzhen Institute of Advanced Technology, Chinese Academy of Science, Shenzhen, China

²OMFS-IMPACT, Department of Imaging and Pathology, KU Leuven, Leuven, Belgium

³Aarhus Institute of Advanced Studies, Aarhus University, Denmark

⁴School of Health Sciences, University of Sydney, Sydney, Australia

⁵Department of Medical Physics, Westmead Hospital, Westmead, Australia

⁶Medical Imaging Research Center and Department of Nuclear Medicine, KU Leuven, Leuven, Belgium

Abstract. Patient movement affects image quality in oral and maxillofacial cone-beam CT imaging. While many efforts are made to minimize the possibility of motion during a scan, relatively little attention has been given to motion correction after the acquisition. We propose a novel method which can improve the image quality after an oral and maxillofacial scan. The proposed method is based on our previous work and is a retrospective motion estimation and motion compensation (ME/MC) approach that iteratively estimates and compensates for rigid pose change over time. During motion estimation, image update and motion update are performed alternately in a multi-resolution scheme to obtain the motion. We propose to use a feature-based motion update and patch-based image update in the iterative estimation process, to alleviate the effect of limited scan field-of-view on estimation. During motion compensation, a fine-resolution image reconstruction was performed with compensation for the estimated motion. The proposed ME/MC method was evaluated with simulations, phantom and patient studies. Two experts in dentomaxillofacial radiology assessed the diagnostic importance of the resulting motion artifact suppression. The quality of the reconstructed images was improved after motion compensation, and most of the image artifacts were eliminated. Quantitative analysis by comparison to a reference image and by calculation of a sharpness metric agreed with the qualitative observation. The results are promising, and further evaluation is required to assess the clinical value of the proposed method.

Keywords: oral and maxillofacial imaging, motion estimation, motion compensation, cone-beam Computed Tomography (CBCT)

1. Introduction

Patient motion is one of the main causes of image artifacts in oral and maxillofacial imaging. Unlike clinical helical CT, dental imaging uses cone-beam CT (CBCT) most commonly, which typically requires longer scan time and thus has a higher probability of patient movement. It has been shown that, due to fear of the tube/detector movement, some patients, particular children, move their head during a scan [1, 2]. Also, it is difficult to prevent aged patients with brain disorders such as Parkinson’s disease from moving [3]. In a study, it was found that about 18% of the pediatric patients and 24% of aged subjects were likely to move sufficiently to induce motion artifacts in the reconstructed images [4].

The motion artifacts in the reconstructed image can degrade the image quality, adversely affecting diagnosis or treatment planning [5]. Even when the motion is too small to be noticed directly, it may still cause a significant resolution loss in the reconstructed image. In such cases, motion, rather than the resolution of the flat panel detector, may become the limiting factor of the image resolution. Further problems arise when metal implants are present together with motion [6]. To obtain an image without artifacts, a retake of the scan may be required, which inevitably causes additional radiation dose to the subject.

Many efforts have been made to prevent patient motion. It is often suggested an examiner should instruct the subjects to close their eyes and be prepared for the scanner movement prior to the acquisition. Patients are sometimes immobilized using a head strap, and/or a chin holder. However, solely relying on the hardware to fix the head may not prevent all potential movements [7]. There are several existing retrospective correction methods focusing on general CBCT applications. Wicklein et al [8], Sisniega et al [9], Jang et al. [10] and Bruder et al. [11] minimize an image-based cost function (which essentially detects motion artifacts) to estimate the motion of extremities, head and lung. Ouadah et al. [12] performed motion correction for head imaging by estimating and compensating for the rigid motion through a 2D-3D registration process, under the precondition that a prior 3D motion-free image of the same subject was available. Ko et al. [13] proposed a generalized CT motion correction framework using a deep neural network, which is capable of real-time correction. However, their model has limited ability to compensate for fast out-of-plane motion and was not applied to any clinical data.

On the other hand, relatively few works have been published on methods, specifically for reducing motion artifacts in reconstructed oral and maxillofacial CBCT images. One recent work demonstrated the potential to track the head movement with the help of an optical camera [14]. Another work aimed to estimate and compensate for the motion retrospectively, by enforcing the locally consistent contours of teeth [15]. We previously described a retrospective method to estimate and compensate for head motion in helical CT imaging, by maximizing the data-consistency between the object and projections [16]. Here, by accounting for the inherent differences between oral and

maxillofacial CBCT and helical CT, we propose a novel method for oral and maxillofacial imaging. The proposed method was evaluated using simulations, a phantom scan, and a patient scan with motion artifacts. Both quantitative and qualitative analysis were performed. Additionally, two experienced dentists trained in dentomaxillofacial radiology further validated the clinical applicability of the proposed method.

2. Materials and Methods




Figure 1. A dental CBCT system usually has a flat detector surface. For each view, the origins of (a) the detector and (b) scanner coordinate systems coincide. R denotes rotation, T denotes translation.

Let us first define the coordinate systems of a virtual CBCT imaging system in Figure 1. We define two coordinate systems. One is the detector coordinate system (Figure 1a), which is fixed with respect to the rotating source-detector pair, and its z-axis coincides with scanner rotation axis. Another is the world coordinate system (Figure 1b), which is fixed with respect to the scanner. For one projection view, we may define a rigid transformation in the scanner system as:

$$S_{\text{world}} = [R_t, R_s, R_c, T_x, T_y, T_z]^T \quad (1)$$

which can be mapped to the detector system as:

$$S_{\text{detector}} = [R_u, R_v, R_z, T_u, T_v, T_z]^T \quad (2)$$

where T_u and T_v are translations tangent and orthogonal to the detector. The aim is first to estimate the pose of the head at each acquired CBCT projection, which we here refer to as motion estimation (ME); then compensate for the estimated motion during the final image reconstruction process, which we refer to as motion compensation (MC). Figure 2 shows the proposed motion estimation and motion compensation (ME/MC) algorithm schematically. ME is an iterative process, in which one iteration contains one motion update and one image update. A similar coordinate system definition and correction scheme to that in Figure 2 was applied to helical scan in [16].

Oral and maxillofacial CBCT typically involves a small field of view and the acquired projections suffer from a strong transaxial truncation, whereas transaxial truncation is usually minor or absent in helical CT. Since ME/MC involves an iterative




Figure 2. Motion Estimation/ Motion Compensation (ME/MC) scheme. ME is an iterative process, of which one iteration involves one motion update step and one image update step. One motion update attempts to adjust the object pose at each view, which minimizes the difference between the measured projections and the re-projected object, with the help of auxiliary projections. One image update step updates the image in the iterative reconstruction process, by incorporating the motion estimates from motion update into the system matrix. MC compensates the estimated motion within a final image reconstruction process.

process with re-projection, insufficient information outside the completely sampled region will induce errors. There is value in minimizing the amount of computation devoted to the background region, which can be done without compromising the accuracy of the motion estimates. More details of the ME and MC implementations are given below.

2.1. Motion update in ME

An initial image is first reconstructed from the measured projections without any motion correction. As in our previous work, we assumed that the rigid pose of the measured object may be different for each projection view. Consequently, a rigid transformation representing the object pose change was estimated for every view, by a 2D-3D registration process. Let r_0 be the current estimate of the one of these parameters (a rotation or translation), r is the value to be estimated for that parameter. Assuming that the actual change in the pose parameter represented by $\hat{r}$ is small, the derivative of projection with respect to r can be approximated as a finite difference of the intensities. For that purpose, we introduce auxiliary projections which are derived by reprojecting the current image estimate with a small change Δr applied to one of the motion parameters. Auxiliary projections are computed to construct the least squares

minimization problem at one view for given iteration:

$$\hat{r} = \underset{r}{\operatorname{argmin}} ||\Delta r(F_{\text{meas}} - F_{\text{reproj}}) - r(F_{\text{incre}} - F_{\text{reproj}})||^2 \quad (3)$$

where Δr is induced increment in r , F_{meas} is the measured projection, F_{reproj} is the re-projection of the current image estimate, and F_{incre} is the auxiliary projection of the image, computed with the small change Δr applied to r . Eq. (3) enables an approximate analytical calculation of one degree-of-freedom motion parameter at each iteration:

$$\hat{r} = \frac{\sum_N (P \cdot Q)}{||\sum_N Q^2||} \Delta r \quad (4)$$

$$\text{where } P = F_{\text{meas}} - F_{\text{reproj}}, \quad Q = F_{\text{incre}} - F_{\text{reproj}}$$

where N is the total number of pixels in projection view θ . We found the translation T_v results in very small magnification of the projection, which can be ignored without adverse effects on final motion corrected image. The newly estimated parameter $\hat{r}$ was added to the current estimate r_0 immediately when estimating the value of the next parameter. The remaining five rigid parameters were estimated in a sequential way at each given projection view (first translation, then rotation). The influence of the induced parameter change on projection was constrained to one degree-of-freedom only each time. This motion estimate for all projection views completes the update of the rigid motion at one iteration, by adding to the current motion estimate. Above process was labeled as “motion update” in Figure 2. Simulations were performed to determine a value of Δr for each degree-of-freedom to be estimated that provided an acceptable balance of speed and accuracy. Excessively large values of Δr were observed to reduce the number of iterations required, thus improving speed, but compromise the accuracy of estimates. Excessively small values increased the number of iterations without improving accuracy. The exact value for each degree-of-freedom used was the largest that did not compromise the estimation accuracy. In this study, Δr was 0.5 mm for all translations and 0.5 degree for all rotations. Above procedure is similar to the one in [16], in which details about the derivation of Eq. (4) can be found.

Our previous method assumed no transaxial truncation in a scan. As explained earlier, this is no longer valid for an oral and maxillofacial scan, as the object outside the region-of-interest (ROI) does not contribute to all acquired views. The truncation artifacts typically result in blurring across the missing projection lines. By focusing on high frequency contents, we expect to reduce the influence of missing projection lines on the motion estimation. Therefore, in this study we did not derive the motion based on the projections themselves but on the high-frequency parts of them.

To do so, we introduced a feature-based motion update. We modified the original motion update by introducing a Laplacian of Gaussian (LOG) operation on all projections (measured projections, re-projections, auxiliary projections). For a given 2D projection view, we applied the LOG operator by applying the convolution kernel:

$$\text{kernel}(x, y) = -\frac{1}{\pi\sigma^4} \left[1 - \frac{x^2 + y^2}{2\sigma^2} \right] \exp \left(-\frac{x^2 + y^2}{2\sigma^2} \right) \quad (5)$$

where σ is the Gaussian smoothing width. A discrete Laplacian kernel that approximates above function was used in the following study, where the 2D window size can be defined. Therefore, the minimization problem in Eq. (3) becomes:

$$\begin{aligned} \hat{r} = \underset{r}{\operatorname{argmin}} & ||\Delta r(\operatorname{LOG}\{P_{\text{meas}}\} - \operatorname{LOG}\{P_{\text{reproj}}\}) \\ & - r(\operatorname{LOG}\{P_{\text{incre}}\} - \operatorname{LOG}\{P_{\text{reproj}}\})||^2 \end{aligned} \quad (6)$$

where LOG denotes the Laplacian of Gaussian operation described above. The rest of the estimation process remained unchanged. In addition, first order gradients, i.e. Sobel and Roberts operators, were applied instead in Eq. (6) to be compared with the LOG filter. The one with the optimal setup that resulted in the best recovered image quality was selected in this study.

2.2. Image update in ME

After obtaining the motion, the image can be updated within an iterative reconstruction process. First let us define the forward model of the acquisition. Assuming monochromatic radiation and ignoring scattering effects, the penalized Poisson log-likelihood function for the attenuation image μ is:

$$L(\mu) = \sum_i y_i \ln \bar{y}_i(\mu) - \bar{y}_i(\mu) - \beta X(\mu) \quad (7)$$

where i is the index of the projection lines, y_i is the measured transmission scan at i , $\bar{y}_i(\mu)$ is the estimated transmission scan at i computed from the attenuation image μ , $X(\mu)$ is the regularization term and β is the weighting factor. We set $\beta = 0$ in this study to ignore the usage of any prior.

By maximizing Eq. (7), one can find the optimal attenuation image iteratively. For example, ignoring the regularization term, the update equation of Maximum Likelihood Transmission-Reconstruction (MLTR) [17] is:

$$\mu_j^{\text{new}} = \mu_j + \frac{\sum_i c_{ij}(\bar{y}_i - y_i)}{\sum_i c_{ij}(\sum_k c_{ik})\bar{y}_i} \quad (8)$$

where μ_j is the linear attenuation coefficient at voxel j , c_{ij} is the intersection length of projection line i with voxel j . Note that we drop the argument of $\bar{y}_i$ for clarity. Compensating for the motion can be done by adjusting the system matrix on-the-fly. Instead of moving the reconstructing image in every view, motion compensation was done by considering a coordinate system fixed to the object and incorporating the motion (now associated to the source-detector pair) into the system matrix. This corresponds to an arbitrary 3D motion of a virtual gantry around the object being scanned, created by the superposition of the inverse of the object motion on the X-ray source trajectory.

To minimize the amount of computation devoted to the background region without compromising the accuracy of motion estimates, we propose to use patch-based reconstruction in image updates. Patch-based reconstruction approaches have been proposed for various applications in CT imaging [18–20]. The idea is that a reconstruction volume can be divided into important patches and not-so-important

patches. For each of these patches a different resolution model can be defined. This allows us to focus the computations more on the interesting regions. We defined two patches in a scan (Figure 3): high-res patch – the patch containing the fully sampled ROI where a smaller voxel size was used; and a low-res patch – the patch comprising the remainder of the object where a coarser resolution model (e.g., voxel size was 4 times larger in all 3 dimensions) was used. The modified update equation of MLTR with the capability to account for patches is:

$$\mu_j^{new} = \mu_j + \frac{a_j \sum_i c_{ij} (\bar{y}_i - y_i)}{\sum_i c_{ij} (\sum_k c_{ik} a_k) \bar{y}_i} \quad (9)$$

where

$$\begin{cases} a_j = 1 & \text{if } j \in \text{patch} \\ a_j = 0 & \text{if } j \notin \text{patch} \end{cases}$$

The patches were updated sequentially, each patch being considered as a group of voxels in a grouped coordinate algorithm. The voxel weighting factor a_j also plays an important role in the convergence of the attenuation values at each voxel. Sequentially updating groups of voxels is known to improve convergence, as the denominator of the update steps will be smaller when the area of the updated patch is smaller [18]. For the one subiteration update in MLTR, we have two consecutive steps (Figure 3): a first update performed only on the high-res patch, and a sequential update performed on low-res patch. MLTR performed such alternate updates for two patches to complete the image update in ME. For motion update, the re-projection process involved the forward projections with different resolutions for different patches, and combination of these projections into one projection.

2.3. Multi-resolution acceleration in ME

The motion update and image update were alternated to increase the likelihood, and the iterations were stopped when the summation of projection errors did not change by more than 2%. These errors between the reprojected and measured data were calculated over all projection views. A multi-resolution approach was applied to accelerate the computation. For example, the starting image resolution level was $4 \times 4 \times 4$, i.e. a down-sampling factor of four was applied in all directions. There was a resampling with a factor of two between adjacent levels. So the image resolution at next level would be $2 \times 2 \times 2$. A similar multi-resolution approach was applied in the projection domain. Since the motion estimate hardly changed during the computations at the finest resolution, we stopped the scheme at the second finest resolution, which significantly reduced the computation. The parameter settings and exact iteration numbers at each resolution level of the multi-resolution scheme for the studies below are provided in Figure 4 and Table 1.




Figure 3. Schematic representation of an iteration of the image update. A first reconstruction update was only performed in the high-res patch, and a sequential reconstruction update was performed in the low-res patch. Then these two patches were combined into a single image. The dashed circle indicates the fully sampled field-of-view (FOV), i.e. high-res patch.




Figure 4. Example multi-resolution update implemented in this study. The number of OSEM-iterations and subsets were applied for the image update at each resolution level are listed. Note that we stop the estimation at the second last level, hence no image update were computed at level 1.

2.4. Final reconstruction in MC

On completion of the iterative ME process, a final reconstruction was performed using the estimated motion. For this reconstruction, a full iterative algorithm can be used that incorporates the motion. Alternatively, to speed up the calculation, an approximate circular Feldkamp-Davis-Kress (FDK) algorithm [21] was implemented, where a first order motion compensation is obtained by taking the motion for each view into account in the back-projection step. This approximation could be applied because the use of physical restraints limited average motion amplitude, which is typically around 0~3 mm [14].

2.5. Quantitative analysis

For the simulation and phantom studies, we computed the root-mean-square error (RMSE), mean structural similarity index measure (MSSIM) [22] and gradient variance (GV) [9, 23] for both uncorrected and corrected images with respect to a reference image, which was reconstructed from a motion-free scan. Taking the motion-corrected image (MCI) as example, RMSE was calculated as:

$$\text{RMSE}(\text{REF}, \text{MCI}) = \sqrt{\frac{1}{N} \sum_j (\mu_{\text{REF},j} - \mu_{\text{MCI},j})^2} \quad (10)$$

where μ_j represented the intensity of voxel j , and N is the total number of voxels. MSSIM index was calculated as:

$$\text{MSSIM}(\text{REF}, \text{MCI}) = \frac{1}{N} \sum_j \text{SSIM}(\mu_{\text{REF},j}, \mu_{\text{MCI},j})$$

$$\text{where } \text{SSIM}(\mu_{\text{REF}}, \mu_{\text{MCI}}) = \frac{(2\bar{\mu}_{\text{REF}}\bar{\mu}_{\text{MC}} + c_1)(2\sigma_{\text{REF-MC}} + c_2)}{(\bar{\mu}_{\text{REF}}^2 + \bar{\mu}_{\text{MC}}^2 + c_1)(\sigma_{\text{REF}}^2 + \sigma_{\text{MC}}^2 + c_2)} \quad (11)$$

$\bar{\mu}$ and σ represent the mean intensity and the variance of the patch, $\sigma_{\text{REF-MC}}$ is the covariance of two patches. The constants $c_1=10^{-4} \text{ cm}^{-2}$ and $c_2=3 \times 10^{-4} \text{ cm}^{-2}$ were used to stabilize the low intensities. GV, which was shown to be sensitive to motion artifacts, was computed as:

$$\text{GV}(\mu) = \sum_j \left(\text{TV}(\mu_j) - \frac{1}{N} \sum_j \text{TV}(\mu_j) \right)^2 \quad (12)$$

where $\text{TV}(\mu_j) = \sqrt{\nabla_x(\mu_j)^2 + \nabla_y(\mu_j)^2 + \nabla_z(\mu_j)^2}, \quad j \in \text{ROI}$

$\text{TV}(\mu_j)$ is the l-2 norm of the image gradients. Since motion would reduce the image resolution hence the sharpness, the corresponding GV is expected to be larger for a motion-free image. As for the patient study, we only computed GV for all images as no reference was available.

Table 1 Implementation details for each experiment.

	Simulation 1	Simulation 2	Simulation 3	Simulation 4	Simulation 5	Phantom	Patient
No. of projections	360	360	360	360	360	512	512
Tube voltage (kVp)	/	/	/	/	/	90	90
Tube current (mA)	/	/	/	/	/	5.0	5.0
Image matrix size at full resolution	512×512 ×256	512×512 ×256	512×512 ×256	512×512 ×256	512×512 ×256	561×561 ×401	481×481 ×481
Voxel size (mm ³)	0.2×0.2 ×0.2	0.2×0.2 ×0.2	0.2×0.2 ×0.2	0.2×0.2 ×0.2	0.2×0.2 ×0.2	0.25×0.25 ×0.25	0.125×0.12× 0.125
Level 3 Iteration no.	8	7	6	3	4	4	5
Level 2 Iteration no.	4	2	4	1	2	2	2
Level 1 Iteration no.	0	0	0	0	0	0	0
Total Iteration no.	12	9	10	4	6	6	7
Computation time (min)	7.5	5.0	6.3	4.1	4.8	36	25

3. Experiments

The proposed ME/MC method was evaluated with simulations, phantom and patient studies. Parameter values used in all studies are listed in Table 1. The details of each experiment are described below.

3.1. Simulations

A digital phantom (Figure 5) was used in all simulations. It was discretized into an image of $512 \times 512 \times 256$ voxels with a voxel size of $0.2 \times 0.2 \times 0.2$ mm³. A detector with 400×300 detector pixels with pixel size of 0.2×0.2 mm² was simulated to create the transaxially truncated projections. The distance between the X-ray source and the detector was 575 mm, the distance between the detector and the rotation center was 216.5 mm.

All simulations are noise-free with a monochromatic source energy of 70 keV. Step-and-shoot acquisition mode was simulated, with a total number of projection views 360 covering one full rotation. The projections were simulated with distance-driven projector [24], while Joseph's ray tracing projector was applied in all reconstructions. We assumed that the motion within one cone-beam projection view is negligible, hence there was no motion simulated within one projection view. Three simulated motion

segments were applied to the digital phantom to compute cone-beam projections. They resembled tremble, slow and abrupt head movement, respectively, as shown in Figure 6a-c. Two segments of head motion from a volunteer recorded over a 5-second period (in a CT scanner but without irradiation) were also applied. Of these two segments one was more abrupt and another was more continuous, see Figure 6d-e.




Figure 5. The digital phantom that used in the simulations. Left: selected axial view; right: coronal and sagittal views.

We compared the image quality in three cases: one reconstructed with motion-free projections as a reference, one with motion-contaminated projections but without any compensation, and one with motion-contaminated projections corrected with the proposed ME/MC method. For reference and uncorrected images, a cone-beam circular FDK algorithm was used for reconstruction. The motion corrected image was computed as follows. In one image update, high-res and low-res patches were updated in an interleaved fashion in a subiteration, as in Figure 3. The high-res and low-res patches had resolutions of $0.2 \times 0.2 \times 0.2 \text{ mm}^3$ and $0.8 \times 0.8 \times 0.8 \text{ mm}^3$, respectively. The total number of ME iterations depended on each scan. The multi-resolution scheme was adopted to accelerate the computation. As explained in Section 2.3, the iterative process was stopped when the relative change of the re-projection errors was below a threshold (2%). A final corrected image was obtained with a modified FDK algorithm compensating for the estimated motion, as in Section 2.4.

We compared the proposed method with the method described in [8]. [8] applied autofocus techniques to motion estimation, which has been previously employed in CBCT for correction of geometric misalignment [26]. Motion was estimated by minimizing the motion artifact metric of the image (image entropy). The minimization was performed together with FDK reconstruction in an iterative minimization process, with respect to the number of points being used for down-sampling the trajectory. Simulation studies were performed with the motion as the ground truth was available.

3.2. Phantom Study

A phantom scan was performed on an 3D Accuitomo 170 scanner (Morita Inc., Kyoto, Japan) in the Oral and Maxillofacial Surgery Department of University Hospital of Leuven. The anthropomorphic phantom model was a Sectional Head Phantom SK150




Figure 6. The motion segments used to generate the motion-contaminated projections: (a) tremble motion S1, (b) slow-drift motion S2, (c) abrupt motion occurred in the middle of the scan S3, (d) and (e) recorded motion segments from subject S4 and S5. Rt, Rs, Rc, Tz, Tx, Ty were rotation and translation parameters defined in Figure 1. Details of how (d) and (e) were recorded can be found in [25].

(Phantom Laboratories, Salem, NY, USA). This phantom has an internal air cavity representing the oral, pharynx and trachea anatomy. The scan was a full rotation scan with 512 views. The total scan lasted for 17.5 s using continuous irradiation, with exposure time at each view was ~ 30 ms. Other scan parameters were: detector size 940×748 , detector pixels 0.2×0.2 mm²; scan FOV 140×100 mm²; distance between the X-ray source and the detector 842 mm; distance between the detector and the rotation center 302 mm.

Motion was induced once by pulling a string attached to the phantom in the middle of the scan, while having the phantom centered in the scanner FOV. After starting the acquisition (~ 5 s), the examiner pulled the non-elastic string gently (lasted for ~ 2 sec) from outside the examination room. The maximal induced amplitude was about 10 degree of rotation and 5 mm of translation. After the induced motion, the phantom returned to a new stable position in which it remained relatively stationary for the rest of the scan. The induced motion was expected to create motion artifacts in the reconstructed image. Following the scan with induced movement, a static scan using the same scan parameters but the movement was performed as a reference.

Similar to the simulation study, we aimed to compare the uncorrected, corrected and reference images. An image of the static scan reconstructed using the scanner software was used as the reference image. The size of this image was $561 \times 561 \times 401$ with voxel size $0.25 \times 0.25 \times 0.25$ mm³. A scanner-reconstructed image of the scan with motion was exported as the uncorrected image. The dimensions of this image were the same as the reference image. No motion compensation or post-processing was applied. The motion corrected image was produced as follows. For the moving scan, raw measurement data were first exported. The high-res patch had a resolution of $0.25 \times 0.25 \times 0.25$ mm³, while the low-res patch had $1 \times 1 \times 1$ mm³. Multi-resolution was used to accelerate the iterative ME. MC was performed with a modified high-res FDK algorithm with estimated motion. The voxel size was the same as that of the reference image.

3.3. Patient Study

A 9-year-old child with cleft lip and palate underwent a CBCT scan for the planning of the follow-up treatment. The reconstructed image contained visible motion artifacts. This scan was acquired on the same scanner as the phantom study and had a full rotation acquisition with 512 views, and a scan time of 17.5 s with continuous irradiation, with exposure time at each view was ~ 30 ms. Detector size was 940×748 , detector pixel was 0.2×0.2 mm², scan FOV was 100×50 mm². Other parameters were the same as the ones used in the phantom study. The image size was $481 \times 481 \times 481$, voxel size was $0.125 \times 0.125 \times 0.125$ mm³. The high-res patch had a resolution of $0.125 \times 0.125 \times 0.125$ mm³, while the low-res patch had $0.5 \times 0.5 \times 0.5$ mm³. Iterative ME was performed to estimate the motion. MC was performed in FDK with the motion accounted for during backprojection to obtain the motion-corrected image.

4. Results

4.1. Simulations

The number of iterations and multi-resolution scheme design for each study can be seen in Table 1. Note that both motion-corrected and uncorrected images were first registered to the reference image. This was done because the estimated motion is not always identical to the true motion, resulting in positional and rotational differences with respect to the reference image. The pose of the reconstructed object is arbitrary and roughly corresponds to the average pose during the scan.

Figure 7 shows the images resulting from the simulation experiments. For all simulations with motion from Figure 6, compared to the reference image, the image without any compensation was clearly contaminated by motion artifacts. After applying ME/MC to the measured data, most of the artifacts were suppressed. The double edges and blurring in the uncorrected images mostly disappeared. The overall shapes of the teeth were recovered and the edges were sharper after correction. The corrected images were more similar to the reference image, as indicated by the higher MSSIM (2.41% in average), GV (7.71% in average) and lower RMSE (11.64% in average). There are some residual ghost structures in Figure 7e, which we consider was because the motion estimation cannot fully recover some slow-moving component in Figure 6e. The reconstruction therefore brought structures from other axial position to the display slice position by attempting to explain the suboptimal motion estimate. Figure 8 shows the difference images obtained by subtracting the reference image from corrected and uncorrected images in selected experiments. The difference was much smaller after correction which indicates the recovery of the image quality. We performed convergence analysis for simulation with motion in Figure 6e, for which the residual projection errors at each iteration were plotted and shown in Figure 13.

The selection of the optimal LOG parameters in Eq. (5), i.e., Gaussian smoothing width σ and Laplacian window size was determined based on the experiment in Figure 9, in which the quality index MSSIM, RMSE and GV were optimal for simulations. The final LOG parameters for motion correction were taken as the average values that resulted in optimal quality indices in all simulations. As a result, $\sigma = 1.0$ and window size $= 5 \times 5$ were applied across experiments. Table 2 demonstrates the superiority of the LOG filter over other first order filters. All operator parameters were optimized to achieve their best performance. All three filters enabled better image quality after correction to some extent, while LOG outperformed the other two in terms of MSSIM, RMSE and GV. Figure 14a shows the visual effects of feature-based motion estimation on recovering image quality for the simulation.

We compared the proposed method with the method described in [8]. Simulation studies were performed with the motion in Figure 6 as the ground truth was available. As in Figure 10, all three image quality indices were superior or at least comparable with our proposed method. The processing time to produce an image (with the simulated motion of Figure 6b) was 5 min for our method and 20.5 min for the method to be

compared.




Figure 7. (a)-(e) Axial planes of simulations with motion shown in Figure 6 (each corresponds to S1-S5). (f) reference plane. After correction, most artifacts were successfully suppressed and details in the teeth were recovered. The MSSIM, RMSE ($\times 10^{-2}$) and GV are listed on top of images. The yellow dashed circle in (f) contains the region where GV was calculated. The red dashed circle in (g) contains the FOV. The display window is $[-0.01, 0.05 \text{ cm}^{-1}]$.




Figure 8. Axial planes of differences between (a) the corrected, (b) the uncorrected image and the reference images. Note here we only show results in three experiments, with the presence of tremble (S1), abrupt (S3) and real subject (S4) motion in Figure 6. After motion correction, the image quality was recovered significantly as indicated by the reduced difference from the reference image. The display window is $[-0.05, 0.05 \text{ cm}^{-1}]$.

Table 2 Comparison of applying different filtering on projections for simulations with motions in Figure 6. “None” indicates no filtering was applied. The number in bold indicates the index with best performance. LOG operations on projections enable recovering the best image quality.

	Motion	LOG	Sobel	Robert	None
MSSIM	S1	0.920	0.923	0.913	0.919
	S2	0.926	0.919	0.912	0.918
	S3	0.931	0.928	0.930	0.926
	S4	0.935	0.919	0.934	0.931
	S5	0.924	0.915	0.922	0.920
RMSE ($\times 10^{-2}$)	S1	0.529	0.544	0.551	0.549
	S2	0.714	0.734	0.740	0.746
	S3	0.694	0.729	0.802	0.732
	S4	0.568	0.566	0.612	0.577
	S5	0.551	0.565	0.575	0.568
GV	S1	0.768	0.765	0.725	0.756
	S2	0.819	0.812	0.821	0.814
	S3	0.789	0.787	0.765	0.776
	S4	0.822	0.825	0.801	0.807
	S5	0.874	0.832	0.827	0.835




Figure 9. The optimal parameters for LOG filtering (Gaussian smoothing width σ and Laplacian kernel window size) were found by searching the ones which resulting the largest MSSIM, together with smallest RMSE and largest GV for all simulations. The value (block) for each possible parameter setup is the average over all simulations with motions in Figure 6.

4.2. Phantom Study

Similar to the simulation studies, both motion-corrected and uncorrected images were first registered to the reference image. ME updated the image and the motion alternately, and a total of six iterations were performed (Table 1). Figure 11 shows the motion estimate and selected planes of these images. Compared to the reference image, the image without any compensation was clearly contaminated by motion artifacts. After applying the proposed compensation method, most of the artifacts were suppressed. Quantitative analysis indicated that the correction indeed recovered the image quality. Figure 11 contains profile plots across a tooth that show the local differences. To quantify the change of the sharpness, we also computed the GV in




Figure 10. Comparison of MSSIM, RMSE ($\times 10^{-2}$) and GV between the corrected images produced from (a) proposed method and (b) the ones from a compared method described in Section 2.6. Two methods produced comparable results in terms of the recovered image quality when the motion pattern was relatively abrupt and random, and the proposed method has better results when motion pattern was slow or complex.

three selected ROIs (containing one right molar, left molar and incisor, respectively, see Figure 11) for all images. GV of the teeth was increased after motion compensation (Table 3). The convergence analysis that demonstrates the residual projection error at each iteration is shown in Figure 13. Figure 14b shows the effects of feature-based motion estimation on recovering image quality for the phantom study.

4.3. Patient Study

Because there was no reference (repeat) scan performed in the patient study, we only compared the motion-corrected image with the scanner reconstructed image. ME updated the image and the motion alternately, and a total of seven iterations were performed (Table 1). Figure 12 shows selected axial and coronal planes of the corrected and uncorrected images. Two dentists assessed the reconstructed images. After applying the proposed method, most of the artifacts were suppressed, and certain diagnostic features were more visible, as indicated by the red arrows in Figure 12. Dentists could identify the contact between the front teeth in the axial views after motion compensation but not before; also, structures of teeth could be better differentiated in coronal views which is useful in diagnosis of dental pathology. The image sharpness was also measured in a similar way to the phantom study. In three selected ROIs (containing one right molar, left molar, incisor), GV increased considerably (at least 18.4%) after compensation, which indicates an increase of image sharpness (Table 3). We performed convergence analysis, which assessed the residual projection error at each iteration as shown in Figure 13. Figure 14c shows the effects of feature-based motion estimation on recovering image quality for the patient study.

Table 3 Gradient variance ($\times 10^8$) as index of image sharpness in teeth.

	reference	no correction	after correction (no LOG)	after correction (w/ LOG)
right molar (phantom)	1.723	1.019	1.398	1.515
left molar (phantom)	0.830	0.103	0.478	0.562
incisor (phantom)	1.203	0.607	0.901	0.921
right molar (patient)	/	0.391	0.436	0.463
left molar (patient)	/	0.601	1.098	1.188
incisor (patient)	/	0.221	0.339	0.425




Figure 11. (a) Motion estimate from the phantom study: up are rotations, bottom are translations. (b) 2D axial planes of the reference, uncorrected and corrected images. Note that the “double edge” motion artifacts in the uncorrected image disappeared in the corrected one. MSSIM and RMSE were only calculated within a circular region containing all teeth. Both dentists agreed that better visualization of bony structures was achieved after compensation. The three teeth measured for sharpness are labeled with number. Note that the background grey circle region is outside the FOV (red dashed circle), hence is not of clinical interest. The display window is $[-100, 3000 \text{ HU}]$.




Figure 12. After applying compensation, blurring and shading artifacts were reduced substantially. As for diagnostic difference, dentists could identify the contact between the front teeth (arrow) in the axial views after motion compensation but not before; also, structures of teeth (arrow) could be better differentiated in coronal views which is useful in diagnosis of dental pathology. Three teeth measured with sharpness are labeled with number by their side. The red dashed circle indicates the FOV. The display window is $[-600, 2800\text{HU}]$.




Figure 13. Residual projection errors at each iteration, obtained by summing the absolute differences between the measured and calculated projections for all pixels. Because of the different pattern and degree of the motion in each experiment, different number of iterations may be required for ME. Most of the error was eliminated after the first few iterations, where the resolution level is low. As the iteration/ resolution levels increase, the residual error tends to stabilize.




Figure 14. Comparison of corrected images with and without applying feature-based motion estimation: (a) simulation S5, (b) phantom and (c) patient study. Uncorrected planes are displayed alongside. While corrections without filtering can recover the quality to some extent, LOG operation on projections enable less residual artifacts and sharper details (red arrows).

5. DISCUSSION

Motion has been shown to have big impact on the high-resolution CBCT image quality [4, 5, 27]. In this paper, we proposed a motion estimation and motion compensation approach for oral and maxillofacial CBCT imaging. It tolerates transaxial truncation and only requires the measured raw data. Since no additional measurements are needed, it can be applied retrospectively to motion-contaminated scan data. The proposed method was evaluated in simulation, phantom and patient studies. After compensation, the quality of the reconstructed image was significantly improved. The method would avoid the need for repeat scans when the patient moves, which could significantly reduce the radiation exposure, especially for children who are more likely to move than adults. This is even more important for a child with cleft lip and palate who often needs a series of scans at their early ages to guide the treatments.

Note that in all cases, the motion was assumed and modeled as rigid motion of an object across all projection views. In practice, non-rigid movement of the tissues due to swallowing, or different motion of the upper and lower jaws, might be present

in a scan. How to deal with such movements is beyond the scope of this paper. Also, together with motion, the presence of dental implants and other metallic objects like fillings and crowns could complicate the artifacts in a reconstructed image. A metal artifact correction technique [28, 29] could have to be combined with the proposed motion correction technique to mitigate the motion and metal artifacts. Perhaps the biggest issue is how to separate the compound effects on image quality from motion and metal. For example in [28], a method was proposed in which the motion was first estimated using an image with preliminary metal artifact reduction. Then full metal artifact reduction was performed by modeling the energy model in a polychromatic reconstruction.

The proposed approach worked reasonably well despite severe artifacts outside the high-res patch, where the reconstruction is handicapped by a limited angle problem. It has been previously shown that the high-res patch can in principle be reconstructed exactly if it contains a portion of air background (intensity known to be zero) [30], which is often the case in dental imaging. Moreover, even in the background, spatial frequencies are better reconstructed if they contribute more information to the measured projections (typically edges which are parallel to the projection lines). This implies that the reconstructed low-res patch still contains the relevant edges needed to align measured and forward projected views, which is the essence of the ME/MC scheme.

Unlike a clinical helical CT scan, a typical oral and maxillofacial scan has a small number of views (300~700) acquired in a full or half rotation. Hence the computation requirement to apply motion correction is less demanding. The entire ME/MC process took ~25 mins (20 mins for ME iterations and 5 mins for the MC/ final reconstruction) for the patient scan. Note that only forward and backward projection processes ran on GPU (NVIDIA k40), so the processing time can still be further reduced. Regarding the time reduction of patch reconstruction and multi-resolution, we found that patch reconstruction alone reduced the total computation time to 1/4, while multi-resolution ME reduce the time to 1/10 in a simulation study. The combination of the two accelerations achieved a 40-fold time reduction. Similar results are expected in phantom and patient studies.

There were still some remaining streak artifacts and blurring in the corrected images. These are possibly due to several limitations of the proposed method: 1) scattering and beam hardening effects were not corrected; 2) during the ME iterations, the entire object is not reconstructed exactly due to the truncation, which may reduce the estimation accuracy; 3) we assumed the movement only happens in between views. Because the motion during the acquisition of a single view is not compensated for, fast motions may produce a residual loss of resolution in the corrected images. Given the fact that the lag time in between views was negligible and the exposure time of one projection view in CBCT is ~30 ms, such movement within views might not be negligible. 4) the final FDK reconstruction is not exact when motion compensation is incorporated. However, we considered the overall motion in an oral and maxillofacial scan to be small since physical restraining during the scan already reduce the possibility

of large movements. Therefore, one can use an analytical algorithm (e.g. FDK), with a first-order motion compensation during the back-projection step, without introducing significant artifacts. When motion is large, a dedicated FDK [10, 31, 32] or an iterative reconstruction algorithm may be more appropriate for the final motion-compensated reconstruction and intermediate images in ME.

Variations in scan protocol could affect the motion correction. For example, a small FOV captures less information in the measurement and increases the extent of truncation, which may degrade the results. Offset scan mode is available on some scanners. With this scan mode, a small detector is used to scan a relatively large FOV, at the cost of increased and asymmetrical truncation in each view. These asymmetrically truncated projections may present a challenge to motion estimation. It is expected that a half scan typically suffers more from cone-beam artifacts, but less from motion artifacts due to the shorter scan time, compared with a full rotation scan. Meanwhile, if FDK reconstruction is used, the more consistent data of a short scan will provide reduced susceptibility to motion artifacts than a full scan, in which inconsistencies could strongly impact the image quality.

6. CONCLUSION

In this study, we proposed an approach to suppress the motion artifacts in oral and maxillofacial CBCT imaging. Results from simulations, phantom and patient studies were presented. The improvement in image quality was assessed both quantitatively and qualitatively. Further clinical evaluation with different scan protocol setups is needed.

ACKNOWLEDGEMENTS

The authors want to thank Dr. Koen Michielsens and Dr. Guozhi Zhang for providing the digital phantom and helpful discussion.

References

- [1] E. Y. Keris, Effect of patient anxiety on image motion artefacts in CBCT, *BMC Oral Health* 17 (2017) 1–9. doi:10.1186/s12903-017-0367-4.
URL <https://dx.doi.org/10.1186/s12903-017-0367-4>
- [2] R. Spin-Neto, L. H. Matzen, L. Schropp, E. Gotfredsen, A. Wenzel, Factors affecting patient movement and re-exposure in cone beam computed tomography examination, *Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology* 119 (5) (2015) 572–578. doi:<https://doi.org/10.1016/j.oooo.2015.01.011>.
- [3] K. Donaldson, S. O'Connor, N. Heath, Dental cone beam CT image quality possibly reduced by patient movement, *Dentomaxillofacial Radiology* 42 (2) (2013) 91866873–91866873. doi:10.1259/dmfr/91866873.
URL <https://dx.doi.org/10.1259/dmfr/91866873>
- [4] R. Spin-Neto, A. Wenzel, Patient movement and motion artefacts in cone beam computed tomography of the dentomaxillofacial region: a systematic literature review, *Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology* 121 (2016) 425–433. doi:10.1016/j.oooo.2015.11.019.
URL <https://dx.doi.org/10.1016/j.oooo.2015.11.019>
- [5] R. Pauwels, K. Araki, J. H. Siewerdsen, S. S. Thongvigitmanee, Technical aspects of dental CBCT: state of the art, *Dentomaxillofacial Radiology* 44 (1) (2015) 20140224–20140224. doi:10.1259/dmfr.20140224.
URL <https://dx.doi.org/10.1259/dmfr.20140224>
- [6] C. Nardi, C. Borri, F. Regini, L. Calistri, A. Castellani, C. Lorini, S. Colagrande, Metal and motion artifacts by cone beam computed tomography (CBCT) in dental and maxillofacial study, *La radiologia medica* 120 (7) (2015) 618–626. doi:10.1007/s11547-015-0496-2.
URL <https://dx.doi.org/10.1007/s11547-015-0496-2>
- [7] T. Hanzelka, J. Dusek, F. Ocsek, J. Kucera, J. Sedy, J. Benes, G. Pavlikova, R. Foltan, Movement of the patient and the cone beam computed tomography scanner: objectives and possible solutions, *Oral Surgery, Oral Medicine, Oral Pathology and Oral Radiology* 116 (2013) 769–773. doi:10.1016/j.oooo.2013.08.010.
- [8] J. Wicklein, Y. Kyriakou, W. A. Kalender, H. Kunze, An online motion-and misalignment-correction method for medical flat-detector CT, 86681S). *International Society for Optics and Photonics* 8668 (2013).
- [9] A. Sisniega, J. W. Stayman, J. Yorkston, J. H. Siewerdsen, W. Zbijewski, Motion compensation in extremity cone-beam CT using a penalized image sharpness criterion, *Physics in Medicine and Biology* 62 (9) (2017) 3712–3734. doi:10.1088/1361-6560/aa6869.
URL <https://dx.doi.org/10.1088/1361-6560/aa6869>
- [10] S. Jang, S. Kim, M. Kim, J. B. Ra, Head motion correction based on filtered backprojection for x-ray CT imaging, *Medical Physics* 45 (2) (2018) 589–604. doi:10.1002/mp.12705.
URL <https://dx.doi.org/10.1002/mp.12705>
- [11] H. Bruder, C. Rohkohl, K. Stierstorfer, T. Flohr, Compensation of skull motion and breathing motion in CT using data-based and image-based metrics, respectively, *SPIE Medical Imaging* 9783 (2016) 97831–97831.
- [12] S. Ouadah, M. Jacobson, J. W. Stayman, T. Ehtiati, C. Weiss, J. H. Siewerdsen, Correction of patient motion in cone-beam CT using 3D-2D registration, *Phys Med Biol* 62 (23) (2009) 8813–8831.
- [13] Y. Ko, S. Moon, J. Baek, H. Shim, Rigid and non-rigid motion artifact reduction in X-ray CT using attention module, *Medical Image Analysis* 67 202–202.
- [14] R. Spin-Neto, L. H. Matzen, L. Schropp, E. Gotfredsen, A. Wenzel, Detection of patient movement during CBCT examination using video observation compared with an accelerometer-gyroscope tracking

- system, *Dentomaxillofacial Radiology* 46 (2) (2017) 20160289–20160289. doi:10.1259/dmfr.20160289. URL <https://dx.doi.org/10.1259/dmfr.20160289>
- [15] M. E. Eldib, M. A. Hegazy, M. H. Cho, M. H. Cho, S. Y. Lee, A motion artifact reduction method for dental CT based on subpixel-resolution image registration of projection data, *Computers in Biology and Medicine* 103 (2018) 232–243. doi:10.1016/j.compbiomed.2018.10.028. URL <https://dx.doi.org/10.1016/j.compbiomed.2018.10.028>
- [16] T. Sun, J.-H. Kim, R. Fulton, J. Nuyts, An iterative projection-based motion estimation and compensation scheme for head x-ray CT, *Medical Physics* 43 (10) (2016) 5705–5716. doi:10.1118/1.4963218. URL <https://dx.doi.org/10.1118/1.4963218>
- [17] K. V. Slambrouck, J. Nuyts, Reconstruction Scheme for Accelerated Maximum Likelihood Reconstruction: The Patchwork Structure, *IEEE Transactions on Nuclear Science* 61 (1) (2014) 173–181. doi:10.1109/tns.2013.2287637. URL <https://dx.doi.org/10.1109/tns.2013.2287637>
- [18] K. V. Slambrouck, J. Nuyts, Metal artifact reduction in computed tomography using local models in an image block-iterative scheme, *Medical Physics* 39 (11) (2012) 7080–7093. doi:10.1118/1.4762567. URL <https://dx.doi.org/10.1118/1.4762567>
- [19] J. A. Fessler, E. P. Ficaro, N. H. Clinthorne, K. Lange, Grouped-coordinate ascent algorithms for penalized-likelihood transmission image reconstruction, *IEEE Transactions on Medical Imaging* 16 (2) (1997) 166–175. doi:10.1109/42.563662. URL <https://dx.doi.org/10.1109/42.563662>
- [20] L. Yu, Y. Zou, E. Y. Sidky, C. A. Pelizzari, P. Munro, X. Pan, Region of interest reconstruction from truncated data in circular cone-beam CT, *IEEE Transactions on Medical Imaging* 25 (7) (2006) 869–881. doi:10.1109/tmi.2006.872329. URL <https://dx.doi.org/10.1109/tmi.2006.872329>
- [21] L. A. Feldkamp, L. C. Davis, J. W. Kress, Practical cone-beam algorithm, *Journal of the Optical Society of America A* 1 (6) (1984) 612–612. doi:10.1364/josaa.1.000612. URL <https://dx.doi.org/10.1364/josaa.1.000612>
- [22] Z. Wang, A. C. Bovik, H. R. Sheikh, E. P. Simoncelli, Image Quality Assessment: From Error Visibility to Structural Similarity, *IEEE Transactions on Image Processing* 13 (4) (2004) 600–612. doi:10.1109/tip.2003.819861. URL <https://dx.doi.org/10.1109/tip.2003.819861>
- [23] M. A. Bueno-Ibarra, J. Alvarez-Borrego, L. Acho, C.-S. Mc, Fast autofocus algorithm for automated microscopes, *Opt Eng* 44 (6) (2005) 63601–63601.
- [24] B. D. Man, S. Basu, Distance-driven projection and backprojection in three dimensions, *Physics in Medicine and Biology* 49 (11) (2004) 2463–2475. doi:10.1088/0031-9155/49/11/024. URL <https://dx.doi.org/10.1088/0031-9155/49/11/024>
- [25] J. H. Kim, T. Sun, A. R. Alcheikh, Z. Kuncic, J. Nuyts, R. Fulton, Correction for human head motion in helical x-ray CT, *Physics in Medicine and Biology* 61 (4) (2016) 1416–1438. doi:10.1088/0031-9155/61/4/1416. URL <https://dx.doi.org/10.1088/0031-9155/61/4/1416>
- [26] Y. Kyriakou, R. M. Lapp, L. Hillebrand, D. Ertel, W. A. Kalender, Simultaneous misalignment correction for approximate circular cone-beam computed tomography, *Physics in Medicine and Biology* 53 (22) (2008) 6267–6289. doi:10.1088/0031-9155/53/22/001. URL <https://dx.doi.org/10.1088/0031-9155/53/22/001>
- [27] L. Li, Z. Chen, X. Jin, H. Yu, G. Wang, Experimental measurement of human head motion for high-resolution computed tomography system design, *Optical Engineering* 49 (6) (2010) 63201–63201.
- [28] T. Sun, J. Nuyts, R. Fulton, Simultaneous correction of motion and metal artifacts in head CT scanning, *IEEE Nuclear Science Symposium and Medical Imaging Conference* 2017 1–3.
- [29] A. Hahn, M. Knaup, M. Brehm, S. Sauppe, M. Kachelrieß, Two methods for reducing moving metal artifacts in cone-beam CT, *Medical Physics* 45 (8) (2018) 3671–3680. doi:10.1002/mp.13060. URL <https://dx.doi.org/10.1002/mp.13060>

- [30] M. Defrise, F. Noo, R. Clackdoyle, H. Kudo, Truncated Hilbert transform and image reconstruction from limited tomographic data, *Inverse Problems* 22 (3) (2006) 1037–1053. doi:10.1088/0266-5611/22/3/019.
URL <https://dx.doi.org/10.1088/0266-5611/22/3/019>
- [31] T. Sun, R. Fulton, J. Nuyts, A method to reduce the data-redundancy artifacts for arbitrary source trajectories in CT imaging, *IEEE Nuclear Science Symposium and Medical Imaging Conference* 2018 (2018) 1–3.
- [32] J. Nuyts, R. Fulton, Iterative FDK reconstruction for helical CT of the head with rigid motion compensation, *The 6th International Conference on Image Formation in X-Ray Computed Tomography* (2020) 248–251.